

DVGW CERT Conformity Certificate

DVGW CERT Konformitätsbescheinigung

AS-0630BT0111

Registration Number
Registriernummer

Field of Application <i>Anwendungsbereich</i>	products for safe connection to the water supply <i>Anschluss sichere Produkte für das Trinkwassernetz</i>
Manufacturer <i>Hersteller</i>	Kaltenbach & Voigt GmbH Bismarckring 39, D-88400 Biberach
Distributor <i>Vertreiber</i>	KaVo Dental GmbH Bismarckring 39, D-88400 Biberach
Product Category <i>Produktart</i>	devices connected to drinking water installation: products for medical use (0630)
Product Description <i>Produktbezeichnung</i>	dentist's chair
Model <i>Modell</i>	Dentaleinheit E 30; E 50; E 50 Life; E 70; E 70 Vision; E 80; E 80 Vision; 1058; 1058 Life
Test Reports <i>Prüfberichte</i>	laboratory control test: from 02.10.2018 (TZW) type testing: FA 020/15 from 28.07.2015 (TZW) type testing: FA 002/14 from 16.09.2014 (TZW) type testing: FA 019/12+FA 020/12 from 10.12.2012 (TZW) KTW testing: KA 0105/18 from 21.08.2018 (TZW) hygienic testing: W-298922-18-SI/to from 04.06.2018 (TZW)
Test Basis <i>Prüfgrundlagen</i>	DVGW W 540-VP (01.08.2010) UBA METALLE (21.11.2018) UBA KTW (07.03.2016)

This product may be connected directly to the drinking water installation. The requirements for protection of drinking water against pollutions due to reflux are in compliance with DIN EN 1717.

Date of Expiry / File No. 08.02.2022 / 19-0173-WNV
Ablaufdatum / Aktenzeichen

19.03.2019 Fk A-1/2

Date, Issued by, Sheet, Head of Certification Body
Datum, Bearbeiter, Blatt, Leiter der Zertifizierungsstelle

DVGW CERT GmbH is an accredited body by DAkkS according to DIN EN ISO/IEC 17065:2013 for certification of products for energy and water supply industry.

DVGW CERT GmbH ist von der DAkkS nach DIN EN ISO/IEC 17065:2013 akkreditierte Stelle für die Zertifizierung von Produkten der Energie- und Wasserversorgung.



DVGW CERT GmbH
Zertifizierungsstelle

Josef-Wirmer-Str. 1-3
53123 Bonn

Tel. +49 228 91 88 - 888
Fax +49 228 91 88 - 993

www.dvgw-cert.com
info@dvgw-cert.com

Type <i>Typ</i>	Technical Data <i>Technische Daten</i>	Remarks <i>Bemerkungen</i>
Dentaleinheit E 30	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit E 50	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit E 50 Life	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit E 70	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit E 70 Vision	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit E 80	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit E 80 Vision	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit 1058	type of protection unit: free outlet (type AA; AB)	
Dentaleinheit 1058 Life	type of protection unit: free outlet (type AA; AB)	

Type Variation <i>Ausführungsvariante</i>	Explanations <i>Erläuterungen</i>
Typ AA	free outlet at the utility sink
Typ AB	free outlet at the dentist's chair

Hints of Utilization / Remarks

Verwendungshinweise / Bemerkungen

diameter of connection: DN 15 (G 1/2")

This product fulfills the requirements acc. §17 (6) of the German Drinking Water Ordinance (TrinkwV 2001), dated 09.01.2018.

The required protective device according to DIN EN 1717 is integral part of the tested component.

All materials in contact with drinking water are hygienic harmless up to the protection point.

This DVGW Conformity certificate does not make any statements about the operation of the tested product.